

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
OGM PLUS 00540/CEPPKKE-G1/4i-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5,0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 55 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1001  
Made in Germany by:  
**ECOLAB**<sup>®</sup> Ecoble Engineering GmbH  
Rolfelsenstrasse 7  
83113 Siegsdorf



Year of manufacture: 2024

Anlaufgrenze 6 [l/h]\*  
lower measuring range limit


Impulszahl 35 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024

Prüfer P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number:	280073
Model description:	Oval gear meter PLUS 00540
OGM PLUS 00540/CEPPKKE-G1/4i-99-0005m-P	
Materials of construction:	PVC, EPDM, PEEK, ceramic
Signal output (standard):	5.0 ml/pulse, 200 pulses/l
Max. pump size:	54 / 95 l/h (50/60Hz)
Semiconductor junction:	PNP
Serial number:	731934 / 11224 / 1002
Year of manufacture:	2024
<b>ECOLAB</b> <sup>®</sup>	Made in Germany by: Ecolab Engineering GmbH Patteleustraße 7 83113 Siegsdorf
	

Anlaufgrenze \_\_\_\_\_ 6 [l/h]\*  
lower measuring range limit

Impulszahl \_\_\_\_\_ 33 [n/min]  
pulse rate  
bei \_\_\_\_\_ 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow


Datum \_\_\_\_\_ 22. März 2024

Prüfer \_\_\_\_\_ P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
Model description: OGM PLUS 00540VCEPPKKE-G1/4i-99-0005m-P  
OGM PLUS 00540VCEPPKKE-G1/4i-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 95 l/h (50/80Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1003  
Year of manufacture: 2024

**ECOLAB**<sup>®</sup>  
Made in Germany by:  
Ecolab Engineering GmbH  
Rathlehnstrasse 7  
83313 Steingdorf



Anlaufgrenze \_\_\_\_\_ [l/h]\*  
lower measuring range limit 6

Impulszahl \_\_\_\_\_ [n/min] 35  
pulse rate  
bei \_\_\_\_\_ [l/h]\* 10  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum \_\_\_\_\_ 22. März 2024

Prüfer \_\_\_\_\_ P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Model description: Oval gear meter PLUS 00540  
OGM PLUS 00540/CEPPKE-G/1/4:99-0005m<sup>3</sup>-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 95 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1004  
Year of manufacture: 2024

Made in Germany by:  
Ecolab Engineering GmbH  
Raiffeisenstrasse 7  
83313 Siegsdorf




Anlaufgrenze 6.8 [l/h]\*  
lower measuring range limit

Impulszahl 35 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024  
Prüfer P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number:	280073
Model description:	Oval gear meter PLUS 00540
OGM PLUS 00540/CEPPKKE-G1/4-99-0005m-P	
Materials of construction:	PVC, EPDM, PEEK, ceramic
Signal output (standard):	5,0 ml/pulse, 200 pulses/l
Max. pump size:	54 / 65 l/h (50/60Hz)
Semiconductor Junction:	PNP
Serial number:	731934 / 11224 / 1005
Year of manufacture:	2024
<b>ECOLAB</b> <sup>®</sup>	Made in Germany by: Ecolab Engineering GmbH Ralfelensrass 7 83313 Siegsdorf
	

Anlaufgrenze 6 [l/h]\*  
lower measuring range limit

Impulszahl 35 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow


Datum 22. März 2024

Prüfer P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
OGM PLUS 00540/CEPPKKE-G1/4:99-0005n-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1008

**ECOLAB**  
Made in Germany by:  
Ecolab Engineering GmbH  
Ralfleienstrasse 7  
83313 Siegsdorf



Year of manufacture: 2024

Anlaufgrenze \_\_\_\_\_ [l/h]\*  
lower measuring range limit 5

Impulszahl \_\_\_\_\_ [n/min]  
pulse rate 34  
bei \_\_\_\_\_ [l/h]\*  
at 10

\* = kontinuierlicher Durchfluss / continuous flow


Datum 22. März 2024

Prüfer P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
Model description: OGM PLUS 00540/CEPPKKE-G/1/4/99-0005m-P  
OGM PLUS 00540/CEPPKKE-G/1/4/99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5,0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1007  
Year of manufacture: 2024

Made in Germany by:  
Ecolab Engineering GmbH  
Rathöferstrasse 7  
85315 Stogsdorf



**ECOLAB**

Anlaufgrenze \_\_\_\_\_ [l/h]\*  
lower measuring range limit 6


Impulszahl \_\_\_\_\_ [n/min]  
pulse rate 32  
bei \_\_\_\_\_ [l/h]\*  
at 10

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024

Prüfer P 427

# Prüfprotokoll OGM<sup>plus</sup>

Model number:	280073
Oval gear meter PLUS 00540	
OGM PLUS 00540/CEPPKKE-G1/4I-99-0005m-P	
Materials of construction: PVC, EPDM, PEEK, ceramic	
Signal output (standard): 5.0 ml/pulse, 200 pulses/l	
Max. pump size:	54 / 65 l/h (50/60Hz)
Semiconductor Junction:	PNP
Serial number:	731934 / 11224 / 1008
<b>ECOLAB</b> <sup>®</sup>	Made in Germany by: Ecolab Engineering GmbH Ralfteichstrasse 7 83313 Siegsdorf
Year of manufacture: 2024	

Anlaufgrenze \_\_\_\_\_ [l/h]\*  
lower measuring range limit 6

Impulszahl \_\_\_\_\_ [n/min]  
pulse rate 35  
bei \_\_\_\_\_ [l/h]\*  
at 10

\* = kontinuierlicher Durchfluss / continuous flow

Datum \_\_\_\_\_  
22. März 2024

Prüfer \_\_\_\_\_  
P 421


417101586



# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
Model description: OGM PLUS 00540/CEP/PKE-G1/4-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1009  
Year of manufacture: 2024

Made in Germany by:  
Ecolab Engineering GmbH  
Raiffeisenstrasse 7  
83313 Siegsdorf



Anlaufgrenze \_\_\_\_\_ [l/h]\*

lower measuring range limit

Impulszahl \_\_\_\_\_ [n/min]

pulse rate

bei \_\_\_\_\_ [l/h]\*

at

\* = kontinuierlicher Durchfluss / continuous flow

Datum \_\_\_\_\_ 22. März 2024


Prüfer \_\_\_\_\_ P 421

417101586

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Model description: Oval gear meter PLUS 00540  
OGM PLUS 00540VCEPPKKE-G1/4i-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1010  
Year of manufacture: 2024

**ECOLAB**<sup>®</sup>  
Made in Germany by  
Ecolab Engineering GmbH  
Rathleisenstrasse 7  
83313 Siegsdorf



Anlaufgrenze 6 [l/h]\*

lower measuring range limit

Impulszahl 33 [n/min]

pulse rate

bei 10 [l/h]\*

at

\* = kontinuierlicher Durchfluss / continuous flow


Datum 22. März 2024

Prüfer P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Model description: Oval gear meter PLUS 00540  
OGM PLUS 00540/CEP/PKKE-G1/4i-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5,0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1011  
Year of manufacture: 2024

**ECOLAB**<sup>®</sup>  
Made in Germany by:  
Ecolab Engineering GmbH  
Rahlfeldstrasse 7  
83313 Siegsdorf



Anlaufgrenze 6 [l/h]\*  
lower measuring range limit


Impulszahl 35 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024

Prüfer P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Model description: Oval gear meter PLUS 00540  
OGM PLUS 00540/CERP/KKE-G/1/4/99-0005h-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/80Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1012  
**ECOLAB**<sup>®</sup> Made in Germany by:  
Ecolab Engineering GmbH  
Ralflehenstrasse 7  
83313 Siegsdorf  
Year of manufacture: 2024 

Anlaufgrenze \_\_\_\_\_ 6 [l/h]\*  
lower measuring range limit

Impulszahl \_\_\_\_\_ 34 [n/min]  
pulse rate

bei \_\_\_\_\_ 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow


Datum \_\_\_\_\_ 22. März 2024

Prüfer \_\_\_\_\_ P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Model description: Oval gear meter PLUS 00540  
OGM PLUS 00540/CEPPKKE-G/1/4/99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1013  
Year of manufacture: 2024

Made in Germany by:  
Ecolab Engineering GmbH  
Raiffeisenstrasse 7  
83313 Siegsdorf



Anlaufgrenze 6 [l/h]\*  
lower measuring range limit

Impulszahl 35 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at


\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024  
Prüfer P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
OGM PLUS 00540/CEPPKKE-G1/4i-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5,0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor Junction: PNP  
Serial number: 731934 / 11224 / 1014

Made in Germany by:  
Ecolab Engineering GmbH  
Ralfleitenstrasse 7  
83313 Steglitzdorf



Year of manufacture: 2024

Anlaufgrenze 6 [l/h]\*  
lower measuring range limit

Impulszahl 32 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024


Prüfer P 421

417101586

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
Model description: OGMPKKE-G1/4-69-0005m-P  
OGM PLUS 00540  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
731934 / 11224 / 1015  
Serial number:  
**ECOLAB**  
Year of manufacture: 2024

Made in Germany by:  
Ecolab Engineering GmbH  
Hallestraße 7  
83313 Stegastorf



Anlaufgrenze 6 [l/h]\*  
lower measuring range limit

Impulszahl 34 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024

Prüfer P 427

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
Model description: OGM PLUS 00540/CEPPKKE-G/1/4i-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor Junction: PNP  
Serial number: 731934 / 11224 / 1016  
**ECOLAB**  
Year of manufacture: 2024

Made in Germany by:  
Ecolab Engineering GmbH  
Ralfleisenstrasse 7  
83133 Siegsdorf



Anlaufgrenze \_\_\_\_\_ 6 [l/h]\*  
lower measuring range limit

Impulszahl \_\_\_\_\_ 34 [n/min]  
pulse rate  
bei \_\_\_\_\_ 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum \_\_\_\_\_ 22. März 2024

Prüfer \_\_\_\_\_ P 427

417101586



# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Model description: Oval gear meter PLUS 00540  
OGM PLUS 00540/CEP/PKKE-G1/4i-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: FNP  
Serial number: 731934 / 11224 / 1017  
Year of manufacture: 2024

**ECOLAB**<sup>®</sup>  
Made in Germany by:  
Ecolab Engineering GmbH  
Fährtenstrasse 7  
85315 Siegsdorf



Anlaufgrenze 6 [l/h]\*  
lower measuring range limit

Impulszahl 34 [n/min]

pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow


Datum 22. März 2024

Prüfer PD 424

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Oval gear meter PLUS 00540  
OGM PLUS 00540/CEPPKKE-G1/4i-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5,0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1018  
**ECOLAB**  
Year of manufacture: 2024

Made in Germany by:  
Ecolab Engineering GmbH  
Raiffeisenstrasse 7  
83313 Siegsdorf



Anlaufgrenze 6 [l/h]\*  
lower measuring range limit

Impulszahl 34 [n/min]  
pulse rate

bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow


Datum 22. März 2024

Prüfer P 421

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
Model description: Oval gear meter PLUS 00540  
OGM PLUS 00540VCEPPKKE-G1/4-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731934 / 11224 / 1019

**ECOLAB<sup>®</sup>**  
Made in Germany by:  
Ecolab Engineering GmbH  
Ralfleisenstrasse 7  
83313 Siegsdorf



Anlaufgrenze 6 [l/h]\*  
lower measuring range limit

Impulszahl 34 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024

Prüfer P 4224

# Prüfprotokoll OGM<sup>plus</sup>

Model number: 280073  
OGM PLUS 00540VCEPPKKE-G1/4;-99-0005m-P  
Materials of construction: PVC, EPDM, PEEK, ceramic  
Signal output (standard): 5.0 ml/pulse, 200 pulses/l  
Max. pump size: 54 / 65 l/h (50/60Hz)  
Semiconductor junction: PNP  
Serial number: 731834 / 11224 / 1020  
Year of manufacture: 2024

Made in Germany by:  
Ecolab Engineering GmbH  
RalfFleckensteinstr. 7  
83313 Siegsdorf



Anlaufgrenze 6 [l/h]\*  
lower measuring range limit

Impulszahl 35 [n/min]  
pulse rate  
bei 10 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22. März 2024

Prüfer P 4 P 421