

Prüfprotokoll OGM^{plus}

Model number: 280003
Model description: Oval gear meter typ 2
OGM 020VCFPPKETI-G1/4I-99-0450-0060
Materials of construction: PVC, FPM, PEEK, ceramic, titan
Signal output (standard): 450 pulses/l
Max. flow rate: 60 l/h

Serial number: 734358 / 12024 / 1001

ECOLAB
Made in Germany by:
Ecolab Engineering GmbH
Raiffelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 11 [l/h]*
lower measuring range limit

Impulszahl 136 [n/min]
pulse rate

bei 18 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 17. Mai 2024

Prüfer P 421

Prüfprotokoll OGM^{plus}

Model number: 280003
Model description: Oval gear meter typ 2
OGM 020VCFPPKETI-G1/4i-99-0450-0060
Materials of construction: PVC, FPM, PEEK, ceramic, titan
Signal output (standard): 450 pulses/l
Max. flow rate: 60 l/h

Serial number: 734358 / 12024 / 1002

ECOLAB

Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 11 [l/h]*
lower measuring range limit

Impulszahl 130 [n/min]
pulse rate

bei 18 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 17. Mai 2024

Prüfer P 421

Prüfprotokoll OGM^{plus}

Model number: 280003
Model description: Oval gear meter typ 2
OGM 020VCFPPKETI-G1/4i-99-0450-0060
Materials of construction: PVC, FPM, PEEK, ceramic, titan
Signal output (standard): 450 pulses/l
Max. flow rate: 60 l/h

Serial number: 734358 / 12024 / 1003

ECOLAB

Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Ralfelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 11 [l/h]*
lower measuring range limit

Impulszahl 137 [n/min]
pulse rate

bei 18 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 17. Mai 2024

Prüfer P 421

Prüfprotokoll OGM^{plus}

Model number:	280003
Model description:	Oval gear meter typ 2
OGM 020VCFPPKKETI-G1/4i-99-0450-0060	
Materials of construction:	PVC, FPM, PEEK, ceramic, titan
Signal output (standard):	450 pulses/l
Max. flow rate:	60 l/h
Serial number:	734358 / 12024 / 1004
ECOLAB	Made in Germany by: Ecolab Engineering GmbH Reiffelsenstrasse 7 83313 Siegsdorf
Year of manufacture: 2024	

Anlaufgrenze 11 [l/h]*
lower measuring range limit

Impulszahl 134 [n/min]
pulse rate

bei 18 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 17. Mai 2024

Prüfer P 421