

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1001
ECOLAB
Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffaisenstrasse
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50.0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1002
ECOLAB
Year of manufacture: 2024
Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50.0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1003

ECOLAB

Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raffelsteinstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1004
ECOLAB
Year of manufacture: 2024
Made in Germany by:
Ecolab Engineering GmbH
Raiffelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50.0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1005
ECOLAB
Year of manufacture: 2024
Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50.0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1006

ECOLAB

Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50.0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1007

ECOLAB

Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1008
ECOLAB
Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1009

ECOLAB

Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Ralfelaenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1010
ECOLAB
Year of manufacture: 2024
Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1011

ECOLAB

Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKE-GS/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1012
ECOLAB
Year of manufacture: 2024
Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1013
ECOLAB
Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1014
ECOLAB
Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4/-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1015

ECOLAB

Year of manufacture: 2024

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50.0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 746529 / 25124 / 1016
ECOLAB
Year of manufacture: 2024
Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
63313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07.01.25

Prüfer P3