

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50.0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1016
ECOLAB
Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Slegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1015

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Räffelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1014

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1013

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Raiffelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1012

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50.0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1011

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Ralfheisenstrasse 7
83313 Slegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number:	280046
Model description:	Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P	
Materials of construction:	PVC, FKM, PEEK, ceramic
Signal output (standard):	50,0 ml/pulse, 20 pulses/l
Max. pump size:	450 / 540 l/h (50/60Hz)
Semiconductor junction:	PNP
Serial number:	748122 / 2725 / 1010
ECOLAB	Made in Germany by: Ecolab Engineering GmbH Raiffelsenstrasse 7 83313 Siegedorf
Year of manufacture: 2025	

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1009

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Ralfelsenstrasse 7
83313 Siedsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1008

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Reifelsenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1007

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Raiffeisenstraße 7
83313 Slegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter: PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1006

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Ralf Heisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1005

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Raifalserstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1004

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Ralfseisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
450 / 540 l/h (50/60Hz)
Max. pump size:
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1003
ECOLAB
Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Ralfseisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
450 / 540 l/h (50/60Hz)
Max. pump size: PNP
Semiconductor junction: 748122 / 2725 / 1002
Serial number:

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Ralfelsenstrasse 7
83313 Slegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3

Prüfprotokoll OGM^{plus}

Model number: 280046
Model description: Oval gear meter PLUS 04500
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P
Materials of construction: PVC, FKM, PEEK, ceramic
Signal output (standard): 50,0 ml/pulse, 20 pulses/l
Max. pump size: 450 / 540 l/h (50/60Hz)
Semiconductor junction: PNP
Serial number: 748122 / 2725 / 1001

ECOLAB

Year of manufacture: 2025

Made in Germany by:
Ecolab Engineering GmbH
Ralf Feisenstrasse 7
83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.02.25

Prüfer P3