Baumer

Passion for Sensors

CERTIFICATE OF CALIBRATION

Type no.:	AFI4-5A10.40$6.1310/9316_8179
Serial no.:	U111.27.X-1375-1607
Date:	30-03-2026 08:46:36
Tag:	

Cell Constant	8,78 1/cm
Tolerance	± 1.0 % FSR , 0 … 1 mS/cm to 0 … 500 mS/cm
	± 1.5 % FSR , 0 … 1000 mS/cm
	± 1.5 % FSR , 0 … 500 µS/cm

After finishing the production process, this sensor has been tested to ensure the highest quality. The sensor has been tested in 13 points, the results are printed below.

Measurements:

Reference Resistor [Ohm]	Norminal Conductivity [mS/cm]	Measured Conductivity [mS/cm]	Max Allowed Error [mS/cm]	Error [mS/cm]
96516,95	0,091	0,094	0,008	0,0030
58727,76	0,150	0,153	0,008	0,0039
9965,47	0,881	0,884	0,010	0,0028
1499,68	5,855	5,845	0,100	-0,0094
999,85	8,781	8,785	0,100	0,0037
500,11	17,556	17,564	0,200	0,0076
300,09	29,258	29,262	0,300	0,0042
180,20	48,724	48,742	0,500	0,0182
120,10	73,107	73,157	1,000	0,0499
80,12	109,588	109,511	2,000	-0,0778
50,12	175,171	174,866	2,000	-0,3047
16,21	541,551	542,558	15,000	1,0070
10,08	870,997	870,785	15,000	-0,2121

Calibrated by: Baumer A/S

(Valid without signature /gültig ohne Unterschrift / valable sans signature)

Baumer A/S · Runetoften 19 · DK-8210 Aarhus V. · Denmark

Phone +45 8931 7611 · info.dk@baumer.com · www.baumer.com