Baumer

Passion for Sensors

CERTIFICATE OF CALIBRATION

Inductive Conductivity Transmitter - Conductivity

Customer: ECOLAB Engineering GmbH
Raiffeisenstraße 7
83313 Siegsdorf, Deutschland

Type no: AFI4-5D30.40$6.1113/9316_8179_Q003
Material no: 11272747
Baumer order no: 2474818-20
Customer order no: 4500526851
Serial no: U134.27.X-1383-8762

Range: 0,5 mS - 100 mS

Test equipment:

Yokogawa SC450G-A-A/Q/UM	no: 92NA03280
Yokogawa SC42-EP08/Q	no: N1D501386
Yokogawa SC42-SP34	no: N1D200437
Keithly 2000	no: 63307

This certificate provides traceability of measurement to recognized national/international standards.

Conductivity:

The accuracy specified in the datasheet of AFIx:

- 0 - 1 mS/cm 1%
- 0 - 2 mS/cm 1%
- 0 - 10 mS/cm 1%
- 0 - 50 mS/cm 1%
- 0 - 200 mS/cm 1%

Measurements:

Conductivity	Range[1]	True conductivity[2]	Measured conductivity (display)	Measured current[2)3)] (Current output)	Deviation conductivity (Display)	Deviation current (Current output)		Accuracy Display	Accuracy current output
[mS/cm]	[mS/cm]	[mS/cm]	[mS/cm]	[mA]	[mS/cm]	mA	[mS/cm]	%	%
0,5	0 - 1	0,460	0,457	11,30981	-0,003	-0,046	-0,003	-0,27	-0,29
1	0 - 2	1,065	1,07	12,54332	0,005	0,027	0,003	0,27	0,17
5	0 - 10	5,143	5,14	12,20146	-0,003	-0,028	-0,017	-0,03	-0,17
25	0 - 50	25,179	25,16	12,03817	-0,019	-0,019	-0,060	-0,04	-0,12
100	0 - 200	105,372	105	12,39631	-0,372	-0,033	-0,419	-0,19	-0,21

1) Selected measuring range on AFIx.
2) Calibration performed without temperature compensation.
3) Measured on the output current with a 100 Ohm as load in series with the Ampere meter.

Date: 26-03-2026 Calibrated by: Le Huynh Huu T_{amb}: 23°C

(Valid without signature / gültig ohne Unterschrift)

Baumer A/S · Runetoften 19 · DK-8210 Aarhus V. · Denmark
Phone +45 8931 7611 · info.dk@baumer.com · www.baumer.com