Baumer

Passion for Sensors

CERTIFICATE OF CALIBRATION

Inductive Conductivity Transmitter - Conductivity

Customer: ECOLAB Engineering GmbH
Raiffeisenstraße 7
83313 Siegsdorf, Deutschland

Type no: AFI4-5D30.40$6.1113/9316_8179_Q003
Material no: 11272747
Baumer order no: 2530402-10
Customer order no: 4500533878
Serial no: U342.27.X-1457-4448

Range: 0,5 mS - 100 mS

Test equipment:

Yokogawa SC450G-A-A/Q/UM	no: 92NA03280
Yokogawa SC42-EP08/Q	no: N1D501386
Yokogawa SC42-SP34	no: N1D200437
Keithly 2000	no: 63307

This certificate provides traceability of measurement to recognized national/international standards.

Conductivity:

The accuracy specified in the datasheet of AFIx:

- 0 - 1 mS/cm 1%
- 0 - 2 mS/cm 1%
- 0 - 10 mS/cm 1%
- 0 - 50 mS/cm 1%
- 0 - 200 mS/cm 1%

Measurements:

Conductivity	Range[1]	True conductivity[2]	Measured conductivity (display)	Measured current[2)3)] (Current output)	Deviation conductivity (Display)	Deviation current (Current output)		Accuracy Display	Accuracy current output
[mS/cm]	[mS/cm]	[mS/cm]	[mS/cm]	[mA]	[mS/cm]	mA	[mS/cm]	%	%
0,5	0 - 1	0,482	0,485	11,75294	0,003	0,045	0,003	0,33	0,28
1	0 - 2	1,071	1,083	12,66636	0,012	0,102	0,013	0,62	0,64
5	0 - 10	5,034	5,05	12,06314	0,016	0,009	0,006	0,16	0,06
25	0 - 50	25,179	25,25	12,07632	0,071	0,019	0,059	0,14	0,12
100	0 - 200	104,475	105,1	12,41429	0,625	0,056	0,704	0,31	0,35

1) Selected measuring range on AFIx.
2) Calibration performed without temperature compensation.
3) Measured on the output current with a 100 Ohm as load in series with the Ampere meter.

Date: 20-08-2026 Calibrated by: Le Huynh Huu T_{amb}: 23°C

(Valid without signature / gültig ohne Unterschrift)

Baumer A/S · Runetoften 19 · DK-8210 Aarhus V. · Denmark
Phone +45 8931 7611 · info.dk@baumer.com · www.baumer.com